

RESEARCH

Open Access

Right ventricular stroke volume assessed by pulmonary artery pulse contour analysis

David Berger^{1*} , Jan Hobi¹, Per W. Möller^{1,2}, Matthias Haenggi¹, Jukka Takala¹ and Stephan M. Jakob¹* Correspondence: david.berger@insel.ch¹Department of Intensive Care Medicine, Inselspital, Bern University Hospital, University of Bern, CH-3010 Bern, Switzerland
Full list of author information is available at the end of the article

Abstract

Background: Stroke volume measurement should provide estimates of acute treatment responses. The current pulse contour method estimates left ventricle stroke volume. Heart-lung interactions change right ventricular stroke volume acutely. We investigated the accuracy, precision, and trending abilities of four calibrated stroke volume estimates based on pulmonary artery pulse contour analysis.

Results: Stroke volume was measured in 9 pigs with a pulmonary artery ultrasound flow probe at 5 and 10 cmH₂O of PEEP and three volume states (baseline, bleeding, and retransfusion) and compared against stroke volume estimates of four calibrated pulmonary pulse contour algorithms based on pulse pressure or pressure integration. Bland-Altman comparison with correction for multiple measurements and trend analysis were performed. Heart rate and stroke volumes were 104 ± 24 bpm and 30 ± 12 mL, respectively. The stroke volume estimates had a minimal bias: -0.11 mL (95% CI -0.55 to 0.33) to 0.32 mL (95% CI -0.06 to 0.70). The limits of agreement were -8.0 to 7.8 mL for calibrated pulse pressure to -10.4 to 11.5 mL for time corrected pressure integration, resulting in a percentage error of 36 to 37%. The calibrated pulse pressure method performed best. Changes in stroke volume were trended very well (concordance rates 73–100%, r^2 0.26 to 0.987, for pulse pressure methods and 71–100%, r^2 0.236 to 0.977, for integration methods).

Conclusions: Pulmonary artery pulse contour methods reliably detect acute changes in stroke volume with good accuracy and moderate precision and accurately trend short-term changes in cardiac output over time.

Keywords: Pulse contour analysis, Right ventricle, Stroke volume

Introduction

The monitoring of cardiac output or stroke volume as a cornerstone of perioperative and critical care is recommended for high-risk surgical patients [1] and patients in persistent shock [2]. Thermodilution with the pulmonary artery catheter is considered the clinical reference method. Stroke volume variation (SVV) based on calibrated or uncalibrated arterial pulse contour analyses [3] is commonly used to assess volume responsiveness or preload dependency (which does not equal the need for volume [4, 5]). Increased right ventricular afterload and right ventricular dysfunction may also result in increased SVV and misleadingly suggest volume responsiveness in pulmonary hypertension [6], right



© The Author(s). 2020 **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

heart failure or pulmonary parenchymal diseases [7, 8], and changing left ventricular afterload conditions [9]. In such clinical situations, the use of a pulmonary artery catheter (PAC) facilitates the assessment of right ventricular stroke volume and differentiation of stroke volume variation caused by right ventricular failure or volume dependency. Despite the lack of proven outcome benefits from its use, the PAC provides the unique ability to assess the right heart and pulmonary circulation and to integrate pulmonary gas exchange into hemodynamic assessment [10, 11]. In contrast to pulse contour methods, current thermodilution methods are not feasible for detecting rapid changes in cardiac output due to slow response time [12] and limited repetitions of intermittent boli. A fast-responding cardiac output measurement is necessary for the dynamic assessment of heart-lung interactions and fluid responsiveness with stroke volume variation [5]. Pulmonary artery (PA) pulse contour analysis could provide a method for rapid assessment of changes in right ventricular stroke volume. The right ventricular output is sensitive to changes in afterload and venous return, whereas the elastic properties of the pulmonary circulation may clearly differ from the systemic arterial tree. The specific behavior of pulmonary vascular impedance characteristics [13, 14] may necessitate adapted calculation models. Pulmonary artery pulse contour analysis was investigated in the 1970s [15–18], but not pursued further despite promising results. The reliable method of intermittent PAC thermodilution used in combination with a calibrated PA pulse contour analysis would provide both a reference cardiac output monitor and a fast-response measure of right-sided stroke volume suitable for dynamic assessment of heart-lung-interaction. The aim of this study was to investigate the precision, accuracy, and trending abilities of four simple pulmonary artery pulse contour methods with Bland-Altman analysis and their interdependence on changes in pulmonary artery elastance and changing volume state.

Methods

This analysis was an independent, post hoc sub-study of an experiment on respiratory maneuvers and venous return. Baseline characteristics have been published before [19, 20]. The study was conducted in accordance with the Guide for the Care and Use of Laboratory Animals (National Academy of Sciences, 1996) and Swiss National Guidelines and was approved by the Commission of Animal Experimentation of Canton Bern, Switzerland (approval BE 71/14), and this manuscript adheres to the applicable ARRIVE guidelines. Additional details on anesthesia and the surgical preparation can be found in reference [19]. In brief, 10 pigs (39.1 ± 1.7 kg body weight) were equipped with a transit time ultrasonic flow probe (PAU Series, Transonic Systems, Ithaca, USA) on the pulmonary artery via a sternotomy. Pulmonary artery pressure (PAP) was measured via a catheter surgically inserted into the pulmonary artery (PA). The measurements were taken in closed-chest conditions without suction on the pleural drains. After surgery, a stabilization period followed, in which a bolus of 100 mL hydroxyethylstarch was given, followed by one repeat 100 mL bolus if stroke volume increased $> 10\%$. This is referred to as *baseline* [19].

The experimental protocol consisted of measurements at *PEEP 5 cmH₂O* (volume state like in baseline) and *10 cmH₂O* in order to investigate the effects of increased intrathoracic pressures with increased afterload on the stroke volume estimates. In the second step, stroke volume was altered from *baseline* (at stable PEEP 5 cmH₂O) by stepwise *bleeding* (6 and 3 mL/kg body weight) and *retransfusion* with volume expansion of the shed blood diluted with 1:1 hydroxyethyl starch [19].

Data was recorded in Labview (National Instruments, TX, USA) with a sampling rate of 100 Hz and extracted with custom-made software (Soleasy, Alea Solutions, Zürich, Switzerland).

Tested methods

We tested four, previously published, calibrated pulse contour algorithms for stroke volume determination from the pulmonary artery pressure tracing, identified in a literature review. The stroke volume estimations were based on pressure amplitude or time-pressure curve integration:

1. Calibrated pulse pressure:

$$SV = K_1 \times PP \text{ [15]}$$

2. Calibrated pulse pressure with correction for systole time:

$$SV = K_2 \times PP \times T_s \text{ [15, 21]}$$

3. Pressure integral:

$$SV = K_3 \times \int_{T_0}^{T_e} (P_{PA} - P_{ed}) dt \text{ [16, 22]}$$

4. Pressure integral with time correction:

$$SV = K_4 \times \int_{T_0}^{T_e} (P_{PA} - P_{ed}) dt (1 + T_s / T_D) \text{ [17, 18]}$$

where K_x is a calibration constant for the respective method; PP indicates pulse pressure (or amplitude); P_{PA} and P_{ed} indicate the instantaneous pulmonary artery and diastolic pressure, respectively; T_e is the time point of end-systole (as defined by the occurrence of the dicrotic notch); T_s and T_d indicate the times for systole and diastole, respectively; and SV denotes stroke volume (see e-Figure 1 in the online supplement). For each animal and stable experimental condition, one hundred beats from the pressure trace recordings were analyzed beat-by-beat with visual identification of the end-diastolic pressure and the dicrotic notch. According to the initial descriptions [15–18, 21], calibration constants were calculated with the first six to eight beats within a series, randomly chosen over respiratory cycles, i.e., calibration was performed for individual animals, in all experimental steps, for each of the four methods. We obtained the reference stroke volume for each respective beat by the integration of the systolic portion of the pulmonary artery flow from the ultrasound flow probe.

For immediate trend analysis, three randomly chosen respiratory cycles were chosen and the inspiratory heartbeat compared to the early and late expiratory heartbeats. The precision for each method expressed as the least significant change (i.e., the smallest true change detectable with 95% certainty) was calculated with data from expiratory beats only, as this represents the minimal physiological variation of stroke volume [23].

Pulmonary artery elastance ($E_a = \text{mPAP}/\text{SV}$) was calculated from the reference stroke volumes and the mean pulmonary artery pressure of the given beat as a surrogate of the end-systolic pressure [24].

Statistical analysis

SigmaPlot 12.5 (Systat Software, Germany) and SPSS 25 (IBM Corporation, USA) were used for statistical and graphical analyses. Normal distribution was assessed with the Shapiro-Wilk test to decide between parametric or non-parametric tests. Comparisons between conditions were done with paired Student's *t* test or Wilcoxon signed ranks test for the PEEP states and with one-way repeated measures analysis of variance (ANOVA) with post hoc Bonferroni correction or Friedman's test with post hoc Tukey's correction for the volume states (baseline, bleeding, and retransfusion). Comparisons between experimental conditions and methods were done with two-way repeated measures ANOVA (within-subject factors experimental condition and method). For the ANOVA, sphericity was assessed with Mauchly's test and corrected for with the Greenhouse-Geisser procedure.

Data is given as mean \pm standard deviation or median (range), depending on the normal distribution. Method agreement was assessed with the Bland-Altman analysis of bias (as an indication of accuracy) and precision (by the limits of agreement and the percentage error) [25] with correction for multiple repeated measurements [26] and Pearson's correlation coefficients with linear regressions performed with the least square methods. Changes over time and trending were addressed using the 4-quadrant plot (with percentage concordance calculation [27] and Lin's correlation coefficient [28]) and polar plots [29] by averaging one hundred beats per condition and comparing the inspiratory to the early and late expiratory beats. The zone of exclusion was chosen to be \pm the least significant change, since a change lower than that would not be considered statistically relevant. Multiple linear regression for the computing constants and the method bias was done with standard hemodynamic parameters as predictors.

Results

One animal died before any measurements were taken. The second animal developed ventricular fibrillation at the beginning of *bleeding* precluding further analysis [19]. Overall, 4300 individual beats were analyzed. Baseline hemodynamic data are given in Table 1. Baseline characteristics have been published before [19].

Method comparison

The average reference stroke volume for all conditions combined was 30 ± 12 mL. All pulse contour-based methods for the estimation of stroke volume from the PA pressure tracing had a minimal bias compared to the reference method, ranging from -0.11 mL (95% CI -0.55 to 0.33) to 0.32 mL (95% CI -0.06 to 0.70). Similar biases could be shown for individual animals (e-figures 2, 3, 4, 5 in the online supplement). The limits of agreement ranged from -8.0 to 7.8 mL for the calibrated pulse pressure to -10.4 to 11.1 mL for the time-corrected pressure integral. The correlation between increases in stroke volume and bias were influenced by an outlying animal (Fig. 1). Multiple linear regression did not identify relevant predictors of the slight increase in bias with increasing stroke volume (Tables e1, e2, e3 and e4 in the online supplement).

The percentage error for stroke volumes varied from 26% for pulse pressure to 30% for time-corrected pulse pressure and 37% for the integration methods. Methods did not influence the bias within the PEEP levels ($p = 0.214$ for interaction) but showed an interaction with the volume state ($p < 0.001$, Tables e5 and e6 in the online supplement). Still,

Table 1 Baseline hemodynamics per condition

	PEEP 5 cmH ₂ O	PEEP 10 cmH ₂ O	<i>p</i> value	Baseline	Bleeding	Retransfusion	<i>p</i> value
Heart rate [s ⁻¹]	88 ± 15	95 ± 23	0.25*	100 ± 20	128 ± 26	108 ± 22	< 0.001
Mean arterial pressure [mmHg]	62 ± 9	60 ± 13	0.378	58 ± 10	48 ± 10	63 ± 12	< 0.001
Mean pulmonary artery pressure [mmHg]	19 ± 3	19 ± 4	0.729	19 ± 2	17 ± 3	25 ± 4	< 0.001
Cardiac output [L/min]	2.94 ± 0.62	2.73 ± 0.67	0.084	2.82 ± 0.73	2.45 ± 0.62	3.40 ± 0.54	< 0.001
Stroke volume (mL)	34.8 ± 11.1	31 ± 12.5	< 0.001*	29.4 ± 11.1	20.6 ± 9.9	33.4 ± 10.9	< 0.001 [†]

Part of these data has been previously published [19]

*Signed rank test

[†]Friedman's ANOVA on ranks

zero was included within the 95% confidence intervals of all biases (indicating that there was no significant difference in the bias to a value of zero). The calibration constants for each method did not significantly change between the PEEP levels, while they tended to change with the volume state, mainly in *retransfusion* (Table e7 in the electronic supplement). From the standard hemodynamic parameters, heart rate contributed most to predict changes in the constant in a multiple linear regression, even though at a low level of correlation (Tables e8, e9, e10 and e11 in the online supplement).

The bias varied widely at low PA elastance values, but in both directions, independently of the method used. With increasing dynamic elastance, the precision increased. There is literally no correlation between elastance and bias (Fig. 2).


Least significant change and interchangeability with the reference method

Our flow probe reference method had a least significant change (LSC) of 3.7 mL. The calibrated pulse pressure method had the lowest LSC of 2.7 mL. The time-corrected pulse pressure had a least significant change of 4.4 mL, and for the integration methods, it was 5 mL. Based on the average stroke volume of 30 mL, the LSC of the reference method translates into a precision of 12.3%. This is higher than what the manufacturer reports (10%).

As our percentage error is a combination of the error in the test method and the reference method, we have back-calculated the true percentage error of our methods, based on a reference method precision, as proposed by Montenij et al. [25] and Critchley and Critchley [30]. They are 24% and 27% for the pulse pressure methods and 35% for the integration methods.

Trend analysis

The averaged stroke volumes over one hundred heartbeats trended changes in cardiac output for all investigated methods with high concordance rates (95 to 100%). Correlation coefficients were high with r^2 from 0.969 to 0.978 and all values lay close to the line of identity. Lin's concordance correlation coefficient indicated very high correlation [0.985 (0.98–0.988), Fig. 3]. The exclusion of a central exclusion zone (\pm LSC)




improved all trending parameters (Fig. 3). The angular bias in the polar plots was below 3° for all methods (Fig. 4).

The Bland-Altman analysis for changes showed a minimal bias ranging from 0.03 to 0.06 mL. The limits of agreement increased from −2.7 to 2.8 mL for the pulse pressure method and from −3.2 to 3.1 mL for the methods based on integration. The resulting percentage error for changes for all methods was within a range of 12% of the average stroke volume (e-Figure 6).


For immediate beat-by-beat trending within the respiratory cycle, concordance rates between 71 and 80% was achieved with r^2 from 0.23 to 0.34 (Fig. 5). Lin's concordance correlation coefficient was moderate [0.475 (0.46–0.54)]. The Bland-Altman plot for changes (e-Figure 7) again indicates a small bias, but large limits of agreement.

Discussion

Our study shows that simple methods based on pulse contour analysis of the pulmonary artery pressure trace are capable of estimating right ventricular stroke volume with a small, clinically insignificant bias and trend changes very well. The smallest bias was found for the simplest method, i.e., calibrated pulse pressure. The limits of agreement and the percentage error were at the border of clinical acceptance for stroke volumes [30], but the averaging of one hundred heartbeats rendered an almost perfect ability to trend changes in cardiac output. This could correct for the rather low precision of an




individual stroke volume measurement, where the wide limits of agreement may render point measurements unreliable. Given the heart rate of the animals, this averaging would report fast changes within around 1 min, which is below the time resolution of any thermodilution system, be it intermittent or continuous. Since we also show that the beat-by-beat tracking is—even though acceptable—less favorable than averaging one hundred beats, the optimal running average or filtering rate would still need to be determined and might depend on the clinical scenario. Continuous thermodilution offers practical advantages and is less operator dependent, but also less accurate than intermittent thermodilution [31, 32] and limited by varying time delays of cardiac output results, which may render rapid clinical deteriorations undetectable [12, 33]. A method as ours, based on pulmonary pulse contour could help to circumvent these inherent limitations to the PAC. Systems for the determination of stroke volume variation with pulse contour based on transpulmonary thermodilution, i.e., left-sided or arterial pulse contour analysis, are limited by pulmonary factors. Respiratory rate, tidal volumes, and lung compliance may introduce an unpredictable phase shift in the stroke volume signals between the right and left ventricle. Right ventricular afterload increases may therefore be a reason for left-sided stroke volume variation, even though averaged right and left stroke volumes must equal each other [8]. Measurements upstream of the lung may not be exposed to these issues, since the right heart directly handles venous return and therefore changes in volume state [19, 20, 34]. Stroke volume variation may




then be of particular interest. Since volume challenges were not part of the study protocol, this question cannot yet be answered. Before a reliable threshold for variations can be defined, the overall measurement performance of a method should be evaluated, which was the aim of this study.

In contrast to previous publications reporting similar strong correlations of estimated and measured stroke volumes [15–18], we assessed the methods based on bias and precision according to Bland and Altman [35, 36] and four-quadrant and polar plots and concordance [27]. Since we have repeated measurements, changes in the variance within the repeated measurements were accounted for [25, 26].

The proportionality between the pulse contour and stroke volume is represented in the calibration constant of the respective methods. Pressure and stroke volume in the pulmonary artery are the components of pulmonary elastance. PEEP may increase right ventricular afterload and pulmonary vascular impedance [7], but we observed no change in calibration constants with the change in PEEP. We found no correlation between the bias and the pulmonary artery elastance, but increasing precision with increasing elastance. This is in contrast to reports suggesting reduced accuracy of thermodilution measurements in pulmonary hypertension [32] and may, if verified in further studies, be a particular strength of our method in patients with acute or chronic pulmonary hypertension. Our calibration constants seemed to change with changing volume state, particularly with *retransfusion*. As a limitation of our retrospective design, the PEEP challenge was unidirectional and of limited intensity. It may introduce spectrum bias. A further limitation is





the reporting of individual stroke volumes from a limited sample of animals. This may lead to an artificial narrowing of the limits of agreement. We have counterbalanced this effect by reporting every animal separately in the online supplement. Sample size calculation for the Bland-Altman analysis has been proposed [37]. This was not possible due to the retrospective study. Since we were particularly interested in trending stroke volume changes, there was no way around reporting individual stroke volumes. As a graphical interpretation model, BA plots may not heavily rely on statistical interferences [25].

Our frequent recalibration may have contributed to the solid trending results. The need for recalibration is an intrinsic problem for any pulse contour system available [3]. As a PAC always offers the possibility for thermodilution and continuous cardiac output measurement systems exist, a self-calibrating system is conceivable and may render short-term changes in stroke volume detectable based on a pulse contour method. The calibrations performed in each condition certainly improved the ability to track changes in SV. This is not currently possible in clinical practice. Combining RV pulse contour SV with recalibration using rapid thermodilution may provide such option and should be assessed in future studies. As we have only used volume changes to change cardiac output, the accuracy of the method may be different for changes in inotropic state or vasoconstriction, and it is an inherent limitation of our study that we do not have thermodilution measurements.

We have used a precise, highly invasive beat-to-beat reference method with a manufacturer-reported precision of $\pm 10\%$. We have found a precision of 12.3%. This reference probe was also used to calculate the calibration constants. It may therefore be criticized that pulmonary pulse contour may work less accurately when calibrated with a thermodilution method, and we may have a coupling between the reference method

and the investigational method. Given our percentage errors of 26 to 37% for our methods under investigation and the reference precision of 12.3%, we cannot claim that our methods would be interchangeable with a highly invasive flow probe. For this to be true, the percentage error should lie below 15% [25, 30]. Clinical interchangeability to an invasive device like a flow probe is not possible anyway. Further studies with thermodilution as a calibration signal are therefore needed. Still, the strength of an intraindividual calibration of the pulse contour signal would also remain in this setting. In the original studies, dye dilution [17, 18], electromagnetic flow meters [16], and intermittent thermodilution with a PAC correlated well with pulse contour methods [15]. None of the original authors described how cardiac output for calibration was measured with respect to the respiratory cycle. This may clearly affect the cardiac output from any dilutional method [38], and the related swings in transpulmonary pressures are a major contributor to changes in right ventricular stroke volume [39]. We have chosen our beats for calibration randomly with respect to airway pressures. The pressure swings induced by mechanical ventilation may however account for the large limits of agreement that we observed. It is therefore of particular importance that the calibration is done with accountance of respiration. Our study is further limited by the small number of animals, the short observation period for trend analysis, and the highly controllable setup of an experimental lab. Still, our results are consistent over the different tested methods and the published literature.

Conclusion

PA pulse contour seems to offer the possibility to detect short-term changes in stroke volume with good accuracy and moderate precision and accurately trend stroke volumes over time. Combined with the detection of waveform changes or pattern recognition, a self-calibrating system for clinical monitoring should be technically feasible.

Supplementary information

Supplementary information accompanies this paper at <https://doi.org/10.1186/s40635-020-00347-7>.

Additional file 1: e-Figure 1. Schematic description of the pulmonary pressure trace (red) and pressure and time components that were used in the four different stroke volume calculations. T_s and T_d denote systolic and diastolic time of the cardiac cycle. $\int T_0 T_e P dt$ indicates the pressure integral over time from T_0 (begin of systole) to T_e (end of systole), whereby T_e minus T_0 equals T_s , the duration of systole.

Additional file 2: e-Figures 2 to 5. Bland-Altman plots for the four methods and all animals with respect to the experimental state. 100 stroke volumes per animal and experimental condition are displayed. The animals are differentiated with a symbol- and color code. The dependencies of bias from the mean of stroke volumes show median r^2 of 0.73 (0.0 to 0.95) for the pulse pressure method, 0.73 (0.0 to 0.94) for the time corrected pulse pressure method, 0.8 (0.01 to 0.95) for the integration method and 0.84 (0.01 to 0.94) for the time corrected pressure integration method.

Additional file 3: e-Figure 6. Bland-Altman plots for changes over experimental conditions. The same data as in Fig. 3 in the main article were used.

Additional file 4: e-Figure 7. Bland-Altman plots for changes over the respiratory cycle. The same data as in Fig. 5 were used.

Additional file 5. Supplementary tables.

Abbreviations

ANOVA: Analysis of variance; E_{dyn} : Dynamic elastance; HR: Heart rate; K_x : Calibration constant; LSC: Least significant change; MAP: Mean arterial pressure; PA: Pulmonary artery; PAC: Pulmonary artery catheter; PAP: Pulmonary artery pressure; P_{ed} : End-diastolic pressure; PEEP: Positive end-expiratory pressure; PP: Pulse pressure; P_{sys} : Systolic pressure; SV: Stroke volume; T_e : End-systole time (dicrotic notch); T_d : Diastolic time; T_s : Systolic time

Acknowledgements

None

Authors' contributions

DB wrote the protocol, participated in the experiments, wrote the initial draft of the manuscript, analyzed the data, performed the statistics, and prepared the graphs. JH participated in the analysis of the raw data and contributed to the writing of the manuscript, in performing the statistics, and in preparing the graphs. PWM participated in performing the experiments and revising the manuscript. MH participated in performing the experiments and revising the manuscript. JT participated in the experiments, helped in writing the protocol, and revised the manuscript. SMJ contributed to the experiments and revised the manuscript. The authors read and approved the final manuscript.

Authors' information

None

Funding

The study was funded by internal departmental funds.

Availability of data and materials

The datasets used and/or analyzed during the current study are available from the corresponding author on reasonable request.

Ethics approval and consent to participate

The study was approved by the animal protection committee of the canton of Bern (BE 71/14).

Consent for publication

Not applicable

Competing interests

The Department of Intensive Care Medicine, University Hospital Bern, has, or has had in the past, research contracts with Orion Corporation, Abbott Nutrition International, B. Braun Medical AG, CSEM SA, Edwards Lifesciences Services GmbH, Kenta Biotech Ltd., Maquet Critical Care AB, and Omnicare Clinical Research AG and research and development/consulting contracts with Edwards Lifesciences SA, Maquet Critical Care AB, and Nestlé. The money was paid into a departmental fund; no author received personal financial gain.

The Department of Intensive Care Medicine has received unrestricted educational grants from the following organizations for organizing a quarterly postgraduate educational symposium, the Berner Forum for Intensive Care (until 2015): Fresenius Kabi, gsk, MSD, Lilly, Baxter, astellas, AstraZeneca, B | Braun, CSL Behring, Maquet, Novartis, Covidien, Nycomed, Pierre Fabre Pharma AG (formerly known as RobaPharm), Pfizer, Orion Pharma, Bard Medica S.A., Abbott AG, and Anandic Medical Systems.

The Department of Intensive Care Medicine has received unrestricted educational grants from the following organizations for organizing bi-annual postgraduate courses in the fields of critical care ultrasound, management of ECMO, and mechanical ventilation: Pierre Fabre Pharma AG (formerly known as RobaPharm), Pfizer AG, Bard Medica S.A., Abbott AG, Anandic Medical Systems, PanGas AG Healthcare, Orion Pharma, Bracco, Edwards Lifesciences AG, Hamilton Medical AG, Fresenius Kabi (Schweiz) AG, Getinge Group Maquet AG, Dräger Schweiz AG, and Teleflex Medical GmbH.

Authors Berger, Hobi, Möller, Haenggi, and Jakob do not have any personal conflicts of interest. Dr. Takala has, after his retirement from the Department of Intensive Care Medicine in August 2018, provided paid consultancy services for the Medical Director and the Director of Technology and Innovation of the Inselspital, Bern University Hospital, and to Nestec SA.

Author details

¹Department of Intensive Care Medicine, Inselspital, Bern University Hospital, University of Bern, CH-3010 Bern, Switzerland. ²Department of Anaesthesiology, Alingsås Hospital, Alingsås, Sweden.

Received: 24 June 2020 Accepted: 29 September 2020

Published online: 07 October 2020

References

1. Vincent JL, Pelosi P, Pearse R, Payen D, Perel A, Hoeft A, Romagnoli S, Ranieri VM, Ichai C, Forget P, Della Rocca G, Rhodes A (2015) Perioperative cardiovascular monitoring of high-risk patients: a consensus of 12. *Crit Care (London, England)* 19:224
2. Cecconi M, De Backer D, Antonelli M, Beale R, Bakker J, Hofer C, Jaeschke R, Mebazaa A, Pinsky MR, Teboul JL, Vincent JL, Rhodes A (2014) Consensus on circulatory shock and hemodynamic monitoring. Task force of the European Society of Intensive Care Medicine. *Intensive Care Med* 40:1795–1815
3. Jozwiak M, Monnet X, Teboul JL (2018) Pressure waveform analysis. *Anesth Analg* 126:1930–1933
4. Takala J (2016) Volume responsive, but does the patient need volume? *Intensive Care Med* 42:1461–1463
5. Pinsky M, Payen D (2005) Functional hemodynamic monitoring. *Crit Care (London, England)* 9:566–572
6. Daudel F, Tüller D, Krähenbühl S, Jakob SM, Takala J (2010) Pulse pressure variation and volume responsiveness during acutely increased pulmonary artery pressure: an experimental study. *Crit Care* 14:R122
7. Grubler MR, Wigger O, Berger D, Blochliger S (2017) Basic concepts of heart-lung interactions during mechanical ventilation. *Swiss Med Wkly* 147:w14491
8. Sondergaard S (2013) Pavane for a pulse pressure variation defunct. *Crit Care* 17:1–6
9. Pinsky MR, Summer WR, Wise RA, Permutt S, Bromberger-Barnea B (1983) Augmentation of cardiac function by elevation of intrathoracic pressure. *J Appl Physiol* 54:950–955

10. Takala J (2007) Hypoxemia due to increased venous admixture: influence of cardiac output on oxygenation. *Intensive Care Med* 33:908–911
11. Richard C, Monnet X, Teboul JL (2011) Pulmonary artery catheter monitoring in 2011. *Curr Opin Crit Care* 17:296–302
12. Reuter DA, Huang C, Edrich T, Shernan SK, Eltzschig HK (2010) Cardiac output monitoring using indicator-dilution techniques: basics, limits, and perspectives. *Anesth Analg* 110:799–811
13. Maggiorini M, Brimiouille S, De Canniere D, Delcroix M, Naeije R (1998) Effects of pulmonary embolism on pulmonary vascular impedance in dogs and minipigs. *J Appl Physiol* (Bethesda, MD: 1985) 84:815–821
14. Brimiouille S, Maggiorini M, Stephanazzi J, Vermeulen F, Lejeune P, Naeije R (1999) Effects of low flow on pulmonary vascular flow-pressure curves and pulmonary vascular impedance. *Cardiovasc Res* 42:183–192
15. Tajimi T, Sunagawa K, Yamada A, Nose Y, Takeshita A, Kikuchi Y, Nakamura M (1983) Evaluation of pulse contour methods in calculating stroke volume from pulmonary artery pressure curve (comparison with aortic pressure curve). *Eur Heart J* 4:502–511
16. DeLoskey AF, Nichols WW, Conti CR, Pepine CJ (1978) Estimation of beat-to-beat stroke volume from the pulmonary arterial pressure contour in man. *Med Biol Eng Comput* 16:707–714
17. Zacharoulis AA, Evans TR, Ziady GM, Coltart DJ, Shillingford JP (1975) Measurement of stroke volume from pulmonary artery pressure record in man. *Br Heart J* 37:20–25
18. Zacharoulis AA, Mills CJ, Gabe IT, Shillingford JP (1974) Estimation of stroke volume from the pulmonary artery pressure record. *Cardiovasc Res* 8:506–516
19. Berger D, Moller PW, Weber A, Bloch A, Bloechlinger S, Haenggi M, Sondergaard S, Jakob SM, Magder S, Takala J (2016) Effect of PEEP, blood volume, and inspiratory hold maneuvers on venous return. *Am J Phys Heart Circ Phys* 311:H794–H806
20. Werner-Moller P, Sondergaard S, Jakob SM, Takala J, Berger D (2019) Effect of volume status on the estimation of mean systemic filling pressure. *J Appl Physiol* (Bethesda, MD: 1985) 126:1503–1513
21. Harley A, Starmer CF, Greenfield JC Jr (1969) Pressure-flow studies in man. An evaluation of the duration of the phases of systole. *J Clin Invest* 48:895–905
22. Wesseling KSN, Nichols W, Weber H, De Wit B, Beneken J (1974) Beat-to-beat cardiac output from the arterial pressure pulse contour. In: Feldman SLJSJ (ed) *Measurement in Anesthesia*. Leiden University Press
23. Odor PM, Bampoe S, Cecconi M (2017) Cardiac output monitoring: validation studies-how results should be presented. *Curr Anesthesiol Rep* 7:410–415
24. Chemla D, Hebert JL, Coirault C, Salmeron S, Zamani K, Lecarpentier Y (1996) Matching diastolic notch and mean pulmonary artery pressures: implications for effective arterial elastance. *Am J Phys* 271:H1287–H1295
25. Montenij LJ, Buhre WF, Jansen JR, Kruitwagen CL, de Waal EE (2016) Methodology of method comparison studies evaluating the validity of cardiac output monitors: a stepwise approach and checklist. *Br J Anaesth* 116:750–758
26. Olofsen E, Dahan A, Borsboom G, Drummond G (2015) Improvements in the application and reporting of advanced Bland-Altman methods of comparison. *J Clin Monit Comput* 29:127–139
27. Saugel B, Grothe O, Wagner JY (2015) Tracking changes in cardiac output: statistical considerations on the 4-quadrant plot and the polar plot methodology. *Anesth Analg* 121:514–524
28. Lin LI (1989) A concordance correlation coefficient to evaluate reproducibility. *Biometrics* 45:255–268
29. Critchley LA, Yang XX, Lee A (2011) Assessment of trending ability of cardiac output monitors by polar plot methodology. *J Cardiothorac Vasc Anesth* 25:536–546
30. Critchley LA, Critchley JA (1999) A meta-analysis of studies using bias and precision statistics to compare cardiac output measurement techniques. *J Clin Monit Comput* 15:85–91
31. Ganter MT, Alhashemi JA, Al-Shabasy AM, Schmid UM, Schott P, Shalabi SA, Badri AM, Hartnack S, Hofer CK (2016) Continuous cardiac output measurement by un-calibrated pulse wave analysis and pulmonary artery catheter in patients with septic shock. *J Clin Monit Comput* 30:13–22
32. Kutter AP, Mosing M, Hartnack S, Raszplewicz J, Renggli M, Mauch JY, Hofer CK (2015) The influence of acute pulmonary hypertension on cardiac output measurements: calibrated pulse contour analysis, transpulmonary and pulmonary artery thermodilution against a modified Fick method in an animal model. *Anesth Analg* 121:99–107
33. Siegel LC, Hennessy MM, Pearl RG (1996) Delayed time response of the continuous cardiac output pulmonary artery catheter. *Anesth Analg* 83:1173–1177
34. Moller PW, Winkler B, Hurni S, Heinisch PP, Bloch A, Sondergaard S, Jakob SM, Takala J, Berger D (2017) Right atrial pressure and venous return during cardiopulmonary bypass. *Am J Phys Heart Circ Phys* 313:H408–h420
35. Bland JM, Altman DG (1986) Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1:307–310
36. Bland JM, Altman DG (1999) Measuring agreement in method comparison studies. *Stat Methods Med Res* 8:135–160
37. Lu M-J, Zhong W-H, Liu Y-X, Miao H-Z, Li Y-C, Ji M-H (2016) Sample size for assessing agreement between two methods of measurement by Bland–Altman method. *Int J Biostat* 12:20150039
38. Jansen JR, Versprille A (1986) Improvement of cardiac output estimation by the thermodilution method during mechanical ventilation. *Intensive Care Med* 12:71–79
39. Pinsky MR (1984) Determinants of pulmonary arterial flow variation during respiration. *J Appl Physiol* 56:1237–1245

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.